



UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE
United States Patent and Trademark Office
Address: COMMISSIONER FOR PATENTS
P.O. Box 1450
Alexandria, Virginia 22313-1450
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/047,896	01/14/2002	Avraham Shekalim	1369-01	5452

7590 04/13/2004

IP Department
Schnader Harrison Segal & Lewis
36th Floor
1600 Market Street
Philadelphia, PA 19103

EXAMINER

ALEXANDER, LYLE

ART UNIT	PAPER NUMBER
----------	--------------

1743

DATE MAILED: 04/13/2004

Please find below and/or attached an Office communication concerning this application or proceeding.

AS

Office Action Summary

Application No.

10/047,896

Applicant(s)

SHEKALIM, AVRAHAM

Examiner

Lyle A Alexander

Art Unit

1743

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If the period for reply specified above is less than thirty (30) days, a reply within the statutory minimum of thirty (30) days will be considered timely.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

Status

- 1) ☐ Responsive to communication(s) filed on ____.
- 2a) ☐ This action is **FINAL**. 2b) ☒ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

Disposition of Claims

- 4) ☒ Claim(s) 1-15 is/are pending in the application.
- 4a) Of the above claim(s) ____ is/are withdrawn from consideration.
- 5) ☐ Claim(s) ____ is/are allowed.
- 6) ☒ Claim(s) 1-15 is/are rejected.
- 7) ☐ Claim(s) ____ is/are objected to.
- 8) ☐ Claim(s) ____ are subject to restriction and/or election requirement.

Application Papers

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☐ The drawing(s) filed on ____ is/are: a) ☐ accepted or b) ☐ objected to by the Examiner.
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

Priority under 35 U.S.C. § 119

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All b) ☐ Some * c) ☐ None of:
1. ☐ Certified copies of the priority documents have been received.
 2. ☐ Certified copies of the priority documents have been received in Application No. ____.
 3. ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

* See the attached detailed Office action for a list of the certified copies not received.

Attachment(s)

- | | |
|---|--|
| 1) <input checked="" type="checkbox"/> Notice of References Cited (PTO-892) | 4) <input type="checkbox"/> Interview Summary (PTO-413)
Paper No(s)/Mail Date. ____ |
| 2) <input type="checkbox"/> Notice of Draftsperson's Patent Drawing Review (PTO-948) | 5) <input type="checkbox"/> Notice of Informal Patent Application (PTO-152) |
| 3) <input type="checkbox"/> Information Disclosure Statement(s) (PTO-1449 or PTO/SB/08)
Paper No(s)/Mail Date ____ | 6) <input type="checkbox"/> Other: ____ |

Claim Rejections - 35 USC § 112

The following is a quotation of the second paragraph of 35 U.S.C. 112:

The specification shall conclude with one or more claims particularly pointing out and distinctly claiming the subject matter which the applicant regards as his invention.

Claims 1-15 are rejected under 35 U.S.C. 112, second paragraph, as being indefinite for failing to particularly point out and distinctly claim the subject matter which applicant regards as the invention.

~~The claims are generally unclear as to the intended~~

Claims 1-10 are rejected under 35 U.S.C. 112, second paragraph, as being incomplete for omitting essential elements, such omission amounting to a gap between the elements. See MPEP § 2172.01. The omitted elements are: means to accomplish the identification that is recited in the preamble, means to collect the sample and what means powers the dynamic needle.

Claims 11-15 are rejected under 35 U.S.C. 112, second paragraph, as being incomplete for omitting essential steps, such omission amounting to a gap between the steps. See MPEP § 2172.01. The omitted steps are: what methods steps are performed to power the dynamic needle and the method correlation between the reaction and identification of the material.

Claims 1-10 are rejected under 35 U.S.C. 112, second paragraph, as being incomplete for omitting essential structural cooperative relationships of elements, such omission amounting to a gap between the necessary structural connections. See MPEP § 2172.01. The omitted structural cooperative relationships are: the relationship between the sleeve and the dynamic needle.

Claim 10 is redundant as it is identical to claim 1 and inherently has the same 112 issues. This claim should be canceled.

Claim 11 line 2 "an outer an outer" is a typographical error and needs correction. The claim is further unclear what is the claimed "shell" and what is intended by "opening an outer shell". The claim is also vague and indefinite where and how the reaction occurs to identify the dangerous material. The specification only states a reaction occurs and does not elaborate any further. Clarification is needed without introducing any new matter.

Claims 3 and 13 are confusing what is intended by a "slider disposed within a slot".

Claim Rejections - 35 USC § 102

The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:

A person shall be entitled to a patent unless –

(b) the invention was patented or described in a printed publication in this or a foreign country or in public use or on sale in this country, more than one year prior to the date of application for patent in the United States.

Claims 1-15 are rejected under 35 U.S.C. 102(b) as being clearly anticipated by Voltz, Warden et al., Levine et al.

The invention is best understood in light of the above 35 USC 112 issues and the specification as an evacuated tube that is punctured to draw up sample into the tube. The tube is puncture again and the sample is contacted with the reagent.

Both Warden et al. and Levine teach collection of blood by vacutainers. The blood sample has been read on the claimed "dangerous material" as blood is a known biohazard. The vacutainer is fitted with needle for collection of the sample by vena puncture which has been read on the claimed puncturing the evacuated tube to draw up the sample. The blood containing tube is again punctured to expose the sample and a reagent.

Voltz teaches a method and apparatus for sampling potentially hazardous device materials. A hollow sampling needle(26) is placed in contact with the sample. A vacuum is drawn and the sample is drawn through the open end(26a) and into the appropriate tubing to supply the sample to a "direct reading instrument". The claimed "fixed needle" has been read on needle(26). The claimed "tube" has been read on the taught tubing attached to the vacuum pump. The claimed "reaction" that identifies the sample on the taught "direct reading instrument".

The prior art made of record and not relied upon is considered pertinent to applicant's disclosure. Shah teach a device for collection of urine using an evacuated container that further contains a reagent for subsequent analysis .

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Lyle A Alexander whose telephone number is 571-272-1254. The examiner can normally be reached on Monday, Wednesday and Friday.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Jill Warden can be reached on 571-272-1267. The fax phone number for the organization where this application or proceeding is assigned is 703-872-9306.

Art Unit: 1743

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free).



Lyle A Alexander
Primary Examiner
Art Unit 1743
